

bastion



Nitrile Soft Blue Powder Free Gloves



bastion

Ultimate solutions in barrier protection

Nitrile Soft Blue Powder Free Gloves

| | |
|------------|--------------------------------------|
| Product: | Nitrile Soft Blue Powder Free Gloves |
| Material: | Nitrile |
| Type: | Non Sterile |
| Design: | Ambidextrous |
| Feature: | Micro Textured Fingers |
| Cuff: | Beaded |
| Usage: | Single Use only |
| Colour: | Sky Blue |
| Packaging: | 100 ± 2 pcs/inner, 10 inners/carton |

Dimensions:

| Size | Palm Width (mm) | MGM Length (mm) | ASTM D3578-05 Length (mm) |
|------|-----------------|-----------------|---------------------------|
| XS | 70 - 79 | Min. 240 | Min. 220 |
| S | 80 - 89 | Min. 240 | Min. 230 |
| M | 90 - 99 | Min. 240 | Min. 230 |
| L | 100 - 109 | Min. 240 | Min. 230 |
| XL | 110 - 119 | Min. 240 | Min. 230 |

Thickness:

| Location of Thickness Measurement | Single Wall (mm) |
|-----------------------------------|------------------|
| Finger | Min. 0.15 |
| Palm | Min. 0.09 |
| Cuff | Min. 0.05 |

Physical Properties:

| | Before Aging | After Aging |
|------------------------|--------------|-------------|
| Tensile Strength (MPa) | Min. 14 | Min. 14 |
| Elongation (%) | Min. 500 | Min. 400 |

| | |
|-----------------------|---|
| Powder Level: | Max. 2mg / dm ² |
| Compliance Standards: | ASTM D6319:00a, EN 455 Part 1,2: 20001, 2 & 3: 2000 |

Inspection Criteria: (Sampling & Inspection ISO 2859)

| | Inspection Level | AQL |
|---------------------|------------------|-----|
| Watertight Test | G1 | 1.5 |
| Major Defects | G1 | 2.5 |
| Minor Defects | G1 | 4.0 |
| Dimensions | S2 | 4.0 |
| Physical Properties | S2 | 4.0 |

Quality Assurance:

ISO 9001: 2000 Certified by SGS (UK) and SIRIM (Malaysia)

ISO 13485: 2003 Certified by SGS (UK)

Directive 89/686/EEC on Personal Protective Equipment Article 11B Certified by SGS (UK)

